

How to use an eakin freeseal®













Prepare the skin

Begin by cleansing the skin around your stoma (1). Ensure that the area for application is clean, dry and free from traces of grease or cream.

Prepare and apply the seal

Warm the seal in your hands for 20 – 30 seconds, until softened, before removing the protective discs.

Option 1 - Gently stretch to fit around the stoma

You may find it easier to apply the eakin **freeseal®** directly to your skin. To apply the seal around your stoma, remove the protective discs from both sides and gently stretch the seal to any shape (2) then place around your stoma and mold into place (3). If you find it easier, the seal may be cut or broken and rejoined by pushing the two ends together.

Once the seal is in place around your stoma ensure that the skin directly around your stoma is not exposed. If it is, you can gently push and mold the seal into place. (4)

Option 2 - Apply directly to the pouching system

Some people find it simpler to apply the seal directly to the pouch baseplate. To apply, remove the protective discs, then mold to the desired shape directly on your baseplate.

Another use of eakin freeseal®

Fill in uneven areas

There may be a need to fill in uneven peristomal skin to help prevent leakage and create a smooth surface for your pouch to stick to. To apply, remove the protective discs from both sides of the seal, break the seal into smaller pieces as required, and mold into uneven skin.

When the eakin **freeseal®** is shaped, press it against your skin. Once the seal is in contact with your skin, you can gently smooth it or mold it further and achieve a fit that suits you (4). Ensure the fit around the stoma is not too tight, as this could prevent normal stoma function.

Attach your pouching system

Two-Piece System: If you have already placed the seal around your stoma, press the baseplate to your skin and over the seal, then apply your pouch to the baseplate. If you've applied the seal directly to the baseplate, apply the baseplate to your skin, then apply your pouch (5).

One-Piece System: If you have already placed the seal around your stoma, apply your pouching system over the seal. If you've applied the seal to your pouch, attach the pouching system to your skin (5).

Always remember to hold your pouch or baseplate in position for a short time, to ensure the seal is secure.

Note: Consult a healthcare professional if you have any skin irritation, redness or discomfort in the area of product application prior to use. If during use you experience any of these symptoms, discontinue use and consult a healthcare professional. **Please see instructions for use for complete instructions.**

convatec me+

For more information contact our me+ team at
1-800-422-8811 or **CIC@convatec.com**

Eakin Healthcare | convatec — forever caring —

© Convatec 2026. The Convatec name, Hug icon and strapline are trademarks of Convatec. Convatec me+ is a trademark of Convatec. ® and ™ indicate trademarks of the Convatec group of companies. eakin **freeseal®** is a trademark of Eakin Healthcare and all copy remains the IP of Eakin Healthcare. AP-78760-USA-ENGU-v1